



DIVISION OF INDUSTRY SERVICES  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
**TTY: Contact Through Relay**

**Scott Walker, Governor**  
**Dave Ross, Secretary**

April 24, 2015

MIELE, INC.  
STEVE POLINSKI  
9 INDEPENDENCE WAY  
PRINCETON NJ 08540

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: MIELE, INC.  
Product Name: MIELE DENTAL DISINFECTOR  
Model Number(s): G7881, G7883, G7883CD, G7804, G7835, G7836, G7831, G7856 AND G7859  
Product File No: 20150077

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of April 2020.

This approval is contingent upon compliance with the following stipulation(s):

- The discharge from this product must discharge to the sanitary drain system through air-gap or air break into a properly sized, trapped and vented, receptor. The piping material shall be stable under the maximum water temperature discharged by these devices.
- No external cross connection control measures are required for these devices because they contain an integral air-gap that conforms to ASME A112.1.2.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter  
Environmental Engineer - Plumbing Product Reviewer  
Department of Safety and Professional Services  
Division of Industry Services  
Bureau of Technical Services  
(608) 267-1401 Phone  
(608) 267-9723 Fax  
glen.schlueter@wi.gov E-mail